



# CERTIFIKAT

No. 31 93 01

## Amalgam Separator A1000B

### Holder / Issued for / Manufacturer

Air Techniques Inc., 1295 Walt Whitman Road, Melville, NY 11747, USA.

### Product name

Amalgam Separator A1000B

### Product description

See appendix

### Certificate

The product described above fulfils the requirements set out in standard SS-EN ISO 11143:1999.

### Marking

Each product that conforms in all respects with the original item certified may display the text 'Type-examined by SP'. In addition, marking shall also show the number(s) of the standard(s), classification/category of the item (if applicable), the number of this certificate and the serial number or equivalent of the item.

### Basis for certification

Test reports No F2 12407 dated 12<sup>th</sup> August 2002, and P3 01270/Rev. Jan 23; 2004, issued by SP Chemistry and Materials Technology.

### Validity

This certificate is valid until not later than 30<sup>th</sup> September 2012.


### Miscellaneous

Other terms and conditions are set out in SP's certification rules for type-examination, SPCR 123. This certificate replaces earlier issues.

Borås, 16<sup>th</sup> October 2007

### SP Technical Research Institute of Sweden Certification

  
Lennart Månsson  
Certification Manager

  
Lennart Aronsson  
Technical Officer

Certificate issued by SP

#### SP Technical Research Institute of Sweden

Postal address	Phone / Fax	Reg.number	E-mail / Internet
SP	+46 10 516 50 00	556464-6874	info@sp.se
Box 857	+46 33 13 55 02		www.sp.se
SE-501 15 Borås			
SWEDEN			

This certificate may not be reproduced other than in full, except with the prior written approval by SP.



Appendix to  
**CERTIFIKAT**  
No. 31 93 01  
dated 16<sup>th</sup> October 2007

The Amalgam separator's exterior is a box with two connections, one inlet and one outlet. Dimensions (length x width x height) approx. 265 x 157 x 180 mm including connections. The interior appears in the drawing "SRAB 99 B", which is a part of the technical documentation.




Photo of actual product: Amalgam Separator A1000B

signed by SP:

**Appendix to certificate issued by SP**

**SP Technical Research Institute of Sweden**

Postal address	Phone / Fax	Reg. number	E-mail / Internet
SP	+46 10 516 50 00	556464-6874	info@sp.se
Box 857	+46 33 13 55 02		www.sp.se
SE-501 15 Borås			
SWEDEN			

This certificate may not be reproduced other than in full, except with the prior written approval by SP.