ACADIA/ACADIA PLUS VISUAL INSPECTION LOG

Serial Number



Best Management Practices (BMP)

- 1 Air Techniques recommends you inspect your Amalgam Separator once a month.
- 2 The collection container must be replaced when **Sediment** (solids) reach the fill line.
- **3** Use Monarch Evacuation System Cleaner when flushing chair-side traps and vacuum lines.
- 4 Purchase a new Collection Cartridge & Recycle Kit from your preferred dealer, PN APREC for ACADIA PLUS and PN A1390 for ACADIA, log onto the Air Techniques portal www.airtechniques.com/login/ to recycle your used container kit and receive your recycling certificate.

As per EPA's BMP please keep records of the following either in physical or electronic form for a minimum of three years:

- Any reports filed
- A visual inspection log
- Documentation of any repair or replacement
- Disposal records
- The manufacturer's operating manual for the current device



Acadia Plus PN APLUS



Acadia PN A1250

Office Name: ,			

Date (MM/DD/Year)	Action/Activity (Routine inspection, change cartridge/filter, ship full cartridge/filter, or repair separator.)	Amalgam Fill %	Person(s) conducting the inspection

For more information please visit:

EPA: https://www.epa.gov/eg/dental-effluent-guidelines

ADA: https://www.ada.org/resources/practice/legal-and-regulatory/amalgam
FDA: https://www.fda.gov/medical-devices/dental-devices/dental-amalgam-fillings

